

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、
当該導入に関して資料等の提供を招請します。

令和7年5月12日

国立大学法人岡山大学長 那須 保友

◎ 調達機関番号 415 ◎ 所在地番号 33

○ 第1号

1 調達内容

(1) 品目分類番号 22、31

(2) 導入計画物品及び数量

ハイブリッド対応血管造影撮影システム
一式

(3) 調達方法 購入等

(4) 導入予定時期

令和7年度第4・四半期以降

(5) 調達に必要とされる基本的な要求要件

A 平面検出器装置及びX線管装置を搭載した血管撮影装置であること。

B 回転DSAの撮影及び3D構築が行えること。

C フラットパネル検出器・X線管部分に安全機能を有し、高精細画像が得られること。

D 血管内治療を支援する機能及び解析機能を有すること。

E 患者及び医療従事者に対する被ばく低減機構を有すること。

F 既存の放射線情報システム（R I S）及び画像ネットワーク（P A C S）と接続が可能であること。

G 線量レポート出力方式がR D S Rに対応していること。

2 資料及びコメントの提供方法 上記1（2）の物品に関する一般的な参考資料及び同（5）の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

(1) 資料等の提供期限 令和7年6月27日17時00分（郵送の場合は必着のこと。）

(2) 提供先 〒700－8558 岡山市北区鹿田町2－5－1 岡山大学病院管理課政府調達担当 平尾詩乃 電話 086－235－7540

3 説明書の交付 本公表に基づき応募する供給者に対して導入説明書を交付する。

(1) 交付期間 令和7年5月12日から令和7年6月27日まで。

(2) 交付場所 上記2（2）に同じ。

4 説明会の開催 本公表に基づく導入説明会を

開催する。

(1) 開催日時 令和7年5月15日10時00分

(2) 開催場所 岡山大学医学部管理棟小会議室

5 その他 この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6 Summary

(1) Classification of the products to be procured : 22, 31

(2) Nature and quantity of the products to be purchased :

Hybrid compatible angiography system 1 Set

(3) Type of the procurement : Purchase

(4) Basic requirements of the procurement :

A The equipment should be an angiography system with a flat panel detector and an X-ray tube.

B The equipment must perform rotation DSA imaging and 3D construction.

C The flat panel detector and X-ray tube must have safety functions. In addition, the equipment must be capable of high-definition images.

D The equipment must have functions to support and analyze endovascular treatment.

E The equipment must have the function of reducing X-rays exposure for patients and medical professionals.

F The equipment must connect our institution's RIS and PACS systems.

G The dose report output method must support RDSR.

(5) Time limit for the submission of the requested material : 17:00 27 June, 2025

(6) Contact point for the notice : HIRAO Shino, Procurement Section, Management Division, Okayama University Hospital, 2 - 5 - 1 Shikata-cho Kita-ku Okayama-shi 700 - 8558 Japan, TEL 086 - 235 - 7540